



NOTE: LIFT THE ANALYZER SYSTEM INTO LOCATION USING LIFTING EYES INSTALLED IN TOP HORIZONTAL UNISTRUT. APPROXIMATE WEIGHT - 600 LBS [272 KG]

REVISION			
REV	DESCRIPTION	DATE	APPROVED
0	FOR INFORMATION ONLY	29FEB12	C. BREWER

FIELD CONNECTION LEGEND		
No.	Service	Connection
1	HEATER POWER	M25
2	ANALYZER POWER	M20
3	ALARM OUTPUTS	M25
4	RS-485 OUTPUT	M25
5	ANALOG OUTPUT	M25
6	INST. AIR INLET (60 PSIG MIN.)	1/4" SS TUBE
7	VAL GAS INLET (30 PSIG)	1/4" SS TUBE
8	SAMPLE SUPPLY	1/4" SS TUBE
9	SAMPLE RETURN (TO FLARE)	1/2" SS TUBE

DIFFERENTIAL PROCESS SYSTEM.
THIS DRAWING IS FOR REFERENCE AND INFORMATIONAL PURPOSES ONLY.

AREA CLASSIFICATION
 ALL ELECTRICAL COMPONENTS ARE CERTIFIED FOR ATEX ZONE 1

DRAWN C. BREWER	DATE 29FEB12	SpectraSensors® 11027 Arrow Route, Rancho Cucamonga, CA 91730 909-948-4100 www.spectrasensors.com	
CHECK		TITLE TYPICAL ANALYZER SYSTEM OUTLINE & MOUNTING DIMENSIONS	
APVD		CUSTOMER P.O. NO.	
SALES ORDER NO.		SIZE D 3XFV0	CAGE CODE EEXD-14-101
CUSTOMER		DRAWING NO. EEXD-14-101	REV 0
CUSTOMER P.O. NO.		AUTOCAD FILE: EEXD-14-101.DWG	SCALE 1/4"=1"
LOCATION			SHEET 1 OF 1

The information contained in this drawing is the sole property of SpectraSensors, Inc. Any reproduction in part or in whole, without the written permission of SpectraSensors, Inc. is prohibited.