



REVISION			
REV	DESCRIPTION	DATE	APPROVED
0	FOR INFORMATION ONLY	28FEB12	C. BREWER

FIELD CONNECTION LEGEND		
No.	Service	Connection
1	HEATER POWER	3/4" CONDUIT
2	ANALYZER POWER	3/4" CONDUIT
3	SIGNAL WIRING	3/4" CONDUIT
4	INST. AIR INLET (60 PSIG MIN.)	1/4" SS TUBE
5	VAL GAS INLET (30 PSIG)	1/4" SS TUBE
6	SAMPLE SUPPLY	1/4" SS TUBE
7	SAMPLE RETURN (TO FLARE)	1/2" SS TUBE

DIFFERENTIAL PROCESS SYSTEM WITH EXTRA FILTRATION.
THIS DRAWING IS FOR REFERENCE AND INFORMATIONAL PURPOSES ONLY.

AREA CLASSIFICATION
 CLASS 1, DIV. 2, GR. B,C&D

DRAWN C. BREWER	DATE 28FEB12	SpectraSensors 11027 Arrow Route, Rancho Cucamonga, CA 91730 909-948-4100 www.spectrasensors.com	
CHECK		TITLE TYPICAL ANALYZER SYSTEM OUTLINE & MOUNTING DIMENSIONS	
SALES ORDER NO.		SIZE D 3XFV0	DRAWING NO. SK2012-12-101
CUSTOMER		AUTOCAD FILE: SK2012-12-101.DWG	SCALE 1/4"=1"
CUSTOMER P.O. NO.		REVISION 0	SHEET 1 OF 1
LOCATION			

The information contained in this drawing is the sole property of SpectraSensors, Inc. Any reproduction in part or in whole, without the written permission of SpectraSensors, Inc. is prohibited.