



1 **EU-TYPE EXAMINATION CERTIFICATE**

2 Equipment intended for use in Potentially Explosive Atmospheres Directive 2014/34/EU

3 Certificate Number: **CSANe 20ATEX1197X** Issue: **0**

4 Equipment: **J22 TDLAS Gas Analyzer, J22 TDLAS Gas Analyzer System with Sample Conditioning System on Panel, J22 TDLAS Gas Analyzer System with Heated Enclosed Sample Conditioning System**

5 Applicant: **SpectraSensors Inc.**

6 Address: **11027 Arrow Route, Rancho Cucamonga, California, 91730, USA**

7 This equipment and any acceptable variation thereto is specified in the schedule to this certificate and the documents therein referred to.

8 CSA Group Netherlands B.V., notified body number 2813 in accordance with Articles 17 and 21 of Directive 2014/34/EU of the European Parliament and of the Council, dated 26 February 2014, certifies that this equipment has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of equipment intended for use in potentially explosive atmospheres given in Annex II to the Directive.

The examination and test results are recorded in the confidential reports listed in Section 14.2.

9 Compliance with the Essential Health and Safety Requirements, with the exception of those listed in the schedule to this certificate, has been assured by compliance with the following documents:

EN IEC 60079-0:2018      EN 60079-1:2014+AC:2018-09      EN 60079-11:2012  
EN 60079-28:2015      EN ISO 80079-36:2016+AC:2019

10 If the sign 'X' is placed after the certificate number, it indicates that the equipment is subject to Specific Conditions of Use identified in the schedule to this certificate.

11 This EU-Type Examination Certificate relates only to the design and construction of the specified equipment. If applicable, further requirements of this Directive apply to the manufacture and supply of this equipment.

12 The marking of the equipment shall include the following:

**J22 TDLAS Gas Analyzer**

**J22 TDLAS Gas Analyzer System with Sample Conditioning System on Panel**



II 2G  
Ex db ib op is IIC T4 Gb  
-20°C ≤ Ta ≤ +60°C



II 2G  
Ex db ib op is h IIC T4 Gb  
-20°C ≤ Ta ≤ +60°C

**J22 TDLAS Gas Analyzer System with Heated Enclosed Sample Conditioning System**



II 2G  
Ex db ib op is h IIC T3 Gb  
-20°C ≤ Ta ≤ +60°C

Project Number    80053042

Signed: J A May

Title: Director of Operations

CSA Group Netherlands B.V.  
Utrechtseweg 310, Building B42,  
6812AR, Netherlands



## SCHEDULE

### EU-TYPE EXAMINATION CERTIFICATE

CSANe 20ATEX1197X

Issue 0

#### 13 DESCRIPTION OF EQUIPMENT

The J22 TDLAS Gas Analyzer is a laser-based gas analyzer that measures the concentration of a compound or “analyte” such as H<sub>2</sub>O. The technology employed is Tunable Diode Laser Absorption Spectroscopy (TDLAS). The measurement output is a volumetric concentration, or ratio of a specific analyte in a gas mixture such as natural gas or air. The volumetric ratio can be converted to other units of measure using internal conversion factors and calculations.

The analyzer is comprised of a sample cell, intrinsically safe optical head and an electronics assembly platform within a pre-certified flameproof enclosure. The cell is a sealed tube through which the gas mixture flows. The cell has a gas inlet and a gas outlet. On one end of the tube is a window through which a beam of infrared laser light travels, which in turn reflects from a mirror. In this arrangement, the gas mixture does not contact the laser or any other optoelectronics. Pressure, and in some cases temperature sensors, are employed in the cell assembly to compensate for the effects of pressure and temperature changes in the gas.

The optical head is mounted on top of the cell and contains the laser, optical detector and a thermoelectric cooler to control the laser temperature. The optical head also contains the optical head electronics which are directly connected to the optoelectronics in the optical head. The optical head electronics board also communicates with the electronics assembly.

The electronics assembly is mounted on top of the optical head, within a flameproof enclosure. The electronics assembly, which can be powered by 100-240 VAC ± 10% or 19.2-28.8 VDC source, contains the sensor electronics which connects to the optical head via the RS232 protocol through a 10-pin ribbon cable assembly. The sensor electronics and the optical head electronics operate on a 30V dc supply using the same 10-pin ribbon cable. The sensor electronics generate the laser drive signal that is sent through the optical head electronics and to the laser in the optical head. Signals from the detectors are amplified by the optical head electronics and sent to the sensor electronics where they are digitized. The sensor electronics process the digital data and sends the gas concentration measurements the electronics display and I/O modules.

The electronics assembly displays the concentration measurement on an LCD display and has a through-the-glass 3-button keypad interface for user input. The flameproof enclosure of the electronics assembly also houses the electrical terminals for field wiring connections. The J22 TDLAS Gas Analyzer comes with various analogue and digital outputs which may be employed in automation or communication systems to deliver its measurements and applicable diagnostic messages and alarms to remote devices. Additionally, the electronics assembly has a webserver which allows the user to interact with the J22 TDLAS Gas Analyzer on a standard web browser using a laptop or tablet.

The J22 TDLAS Gas Analyzer is used as an “extractive” measuring device, where the gas sample is extracted from a vessel or pipeline and transported to the analyzer which may be mounted up to 100 meters from the sample tap point. The J22 TDLAS Gas Analyzer may be configured without sample conditioning. Optional hardware is also available to condition the sample before it enters the gas analyzer. An optional IP66 rated enclosure may be included which surrounds the “cell” and the sample conditioning system. The sample conditioning system (SCS) is required to filter and remove solid and liquid particulate

CSA Group Netherlands B.V.  
Utrechtseweg 310, Building B42,  
6812AR, Netherlands



## SCHEDULE

### EU-TYPE EXAMINATION CERTIFICATE

CSANe 20ATEX1197X  
Issue 0

and to control the flow and pressure of the gas. In some cases, an optional heater (powered independently from the analyzer) is used to control the temperature inside the enclosed SCS. An optional pressure relief valve and purge system are available. The effectiveness of this purge system has not been assessed by CSA.

The J22 TDLAS Gas Analyzer operates at near-atmospheric pressure and at an ambient temperature between -20 to +60 degrees centigrade. After the passing through the J22 TDLAS Gas analyzer, the sample is routed and vented to a safe location in the atmosphere or vented to a flare or other apparatus.

The equipment has been separately tested against the requirements of IEC 60529 and it meets IP66.

#### Model Code Structure

J22 – ABCDEFGHIJKLMNOPQRSTUVWXYZ

A – Approval

BA - ATEX / IECEx Zone 1

B – Analyte

C – Measurement Range

D – Measurement Range 2

E – Stream Composition

F – Venting to

G – Process Wetted Materials

V - 316 Stainless Steel; FKM Seals

H – Supply Parameters

A - 100-240 VAC (50/60 Hz) ± 10%

D – 24 VDC ± 20%

I – Output; Input 1

J – Output; Input 2

K – Output; Input 3

L – Electronics Housing

1 - Coated Copper-Free Aluminum

M – Controller Mounting

N – Sample Conditioning System (SCS)

A - On Panel, Aluminum

B - Enclosed, 304 Stainless Steel

N – None

O – Filtration

P – Sample System Gas Connections

A – Imperial

B – Metric

Q – Pressure Regulation

R – Flow Meter

A - Armored, factory default

B - Armored, Krohne

CSA Group Netherlands B.V.  
Utrechtseweg 310, Building B42,  
6812AR, Netherlands



**SCHEDULE**

**EU-TYPE EXAMINATION CERTIFICATE**

**CSANe 20ATEX1197X  
Issue 0**

- F - Glass Tube, factory default
- K - Glass Tube, Krohne
- N - None
- S – Heating Options
  - 1 - Heated + Heat-Trace Boot, 100 - 240 VAC ± 10%
  - 8 – None
- T – Purge
- U – Operating Language Display
- V – Test/Certificate/Declaration
- W – Marking

Headings without sub-options are not considered critical to the design of the equipment. Where sub options are shown, these are the only options endorsed by CSA.

**14 DESCRIPTIVE DOCUMENTS**

**14.1 Drawings**

Refer to Certificate Annexe.

**14.2 Associated Reports and Certificate History**

Issue	Date	Report number	Comment
0	18 February 2021	R80053042A	The release of the prime certificate.

**15 SPECIFIC CONDITIONS OF USE (denoted by X after the certificate number)**

- 15.1 The flameproof joints of this equipment are other than the minimums specified in IEC/EN 60079-1 and shall not be repaired by the user.
- 15.2 Adhesive labels and the powder coating of models of the equipment with an aluminium enclosure are non-conducting materials and may generate an ignition-capable level of electrostatic discharge under certain extreme conditions. The user should ensure that the Equipment is not installed in a location where it may be subjected to external conditions (such as high-pressure steam) which might cause a build-up of electrostatic charges on these non-conducting surfaces. Additionally, cleaning of the equipment should be done only with a damp cloth.
- 15.3 The optional stainless-steel label tag is not bonded to earth. The maximum average capacitance of the tag determined by measurement is max. 30 pF. This shall be considered by the user to determine suitability of the equipment in a specific application.
- 15.4 For models of the J22 TDLAS Gas Analyzer with SCS mounted within an enclosure, the inner sheath of the supply cable for the heater circuit shall be sheathed with thermoplastic, thermosetting, or elastomeric material. It shall be circular and compact. Any bedding or sheath shall be extruded. Fillers, if any, shall be non-hygroscopic. The minimum length of the cable shall exceed 3 meters.
- 15.5 The temperature of the process medium shall be within the ambient temperature rating of the equipment.

CSA Group Netherlands B.V.  
Utrechtseweg 310, Building B42,  
6812AR, Netherlands



## SCHEDULE

### EU-TYPE EXAMINATION CERTIFICATE

CSANe 20ATEX1197X  
Issue 0

#### 16 ESSENTIAL HEALTH AND SAFETY REQUIREMENTS OF ANNEX II (EHSRs)

The relevant EHSRs that are not addressed by the standards listed in this certificate have been identified and individually assessed in the reports listed in Section 14.2.

#### 17 CONDITIONS OF MANUFACTURE

17.1 The use of this certificate is subject to the Regulations Applicable to Holders of CSA Certificates.

17.2 Holders of EU-Type Examination Certificates are required to comply with the conformity to type requirements defined in Article 13 of Directive 2014/34/EU.

17.3 The equipment covered by this certificate incorporates previously certified devices; it is therefore the responsibility of the manufacturer to continually monitor the status of the certification associated with these devices, and the manufacturer shall inform CSA UK of any modifications of the devices that may impinge upon the explosion safety design of the equipment.

Description	Certificate Number
E+H G305 and G307 Enclosures	IECEX SIR 11.0050U
Intertec SL-***THERM Block Heater	PTB 02 ATEX 1116 X
Intertec TA Thermostat	PTB 03 ATEX 1136 X
Adalet XIHS Terminal Box	DEMKO 12 ATEX 1115099U
CMP Type 737 Adapter/Reducer	CML 18ATEX1320X
CMP A2F Cable Gland	CML 18ATEX1321X IECEX CML 18.0179X
Hazardous Locations Solutions Type N conduit elbow	IECEX SIR 07.0044U
Proline 300/500 Electronics	IECEX CSA 16.0006U
Display Type DP-PA**	IECEX KEM 08.0048X

CSA Group Netherlands B.V.  
Utrechtseweg 310, Building B42,  
6812AR, Netherlands

# Certificate Annexe



Certificate Number: CSANe 20ATEX1197X

Equipment: J22 TDLAS Gas Analyzer, J22 TDLAS Gas Analyzer System with Sample Conditioning System on Panel, J22 TDLAS Gas Analyzer System with Heated Enclosed Sample Conditioning System

Applicant: SpectraSensors Inc.

## Issue 0

Drawing	Sheets	Rev.	Date (Stamp)	Title
490002311	1 to 4	C	16 Feb 21	J22 Analyzer Product Description Document
MPI-100025	1 of 1	A	16 Feb 21	Manufacturing Process Instruction Potting Process RT 622, Exd and CLASS-1 DIVISION-1 XP
EX0900000012	1 to 2	E	16 Feb 21	Housing, Conn Stem, H328
EX0900000025	1 to 2	C	16 Feb 21	Enclosure, Optical Head
EX0900000026	1 of 1	B	16 Feb 21	Cover, Opt Head Rnclosure
EX0900000028	1 of 1	A	16 Feb 21	Hot Plate, TLM
EX0900000029	1 of 1	A	16 Feb 21	Cold Plate, TLM
EX0900000030	1 of 1	A	16 Feb 21	Cover, Opt Block
EX0900000032	1 of 1	A	16 Feb 21	Opt block, WW, DD, Cert
EX1300000038	1 of 1	B	16 Feb 21	Gasket, H328 Collar / Opt Enclsr
EX1300000039	1 of 1	B	16 Feb 21	Gasket, Enclosure Access
EX1700000004	1 to 5	B	16 Feb 21	Label, J22 Rating
EX2100000002	1 to 2	B	16 Feb 21	Bushing + 28 Contacts
EX2100000010	1 to 2	A	16 Feb 21	Conn, 28-Pin, 90V Isolation
EX2900000010	1 to 4	C	16 Feb 21	PCBA, Optical Head Assembly
EX2900000010-00	1 to 3	D	16 Feb 21	BOM, Optical Head Electronics
EX2900000011	1 to 12	A	16 Feb 21	PCB, Optical Head Fabrication Notes
EX2900000012	1 to 9	A	16 Feb 21	Schem, Optical Head Electronics
EX2900000020	1 to 4	D	16 Feb 21	PCBA, ISEM Analog Board Assembly
EX2900000020-00	1 to 3	D	16 Feb 21	BOM, ISEM Analog
EX2900000021	1 to 14	A	16 Feb 21	PCB, ISEM Analog Board Fabrication Notes
EX2900000022	1 to 9	B	16 Feb 21	Schem, ISEM Analog
EX2900000030	1 to 4	C	16 Feb 21	PCBA, ISEM Digital MCU Assembly
EX2900000030-00	1 to 2	C	16 Feb 21	BOM, ISEM Digital MCU Board
EX2900000031	1 to 16	A	16 Feb 21	PCB, ISEM Digital MCU Fabrication Notes
EX2900000032	1 to 17	A	16 Feb 21	Schem, ISEM Digital MCU
EX2900000040	1 to 3	A	16 Feb 21	PCBA, Transition-Laser/Detector Assembly Side-A
EX2900000040-00	1 to 2	A	16 Feb 21	BOM, PCBA Assy, Transition – Laser/Detector
EX2900000041	1 to 10	A	16 Feb 21	PCB, Transition-Laser/Detector Fabrication Notes
EX2900000042	1 to 3	A	16 Feb 21	Schem, Transition – Laser/Detector
EX2900000071	1 to 5	B	16 Feb 21	PCB, TLM Flex Fabrication Notes
EX2900000072	1 to 3	B	16 Feb 21	Schem, TLM Flex
EX2900000081	1 to 7	C	16 Feb 21	PCB, ISEM 28 Flex, Analog Fabrication Notes
EX2900000082	1 to 3	A	16 Feb 21	Schem, ISEM 28 Flex, Analog
EX2900000091	1 to 7	C	16 Feb 21	PCB, ISEM 28 Flex, Digital Fabrication Notes
EX2900000092	1 to 3	A	16 Feb 21	Schem, ISEM 28 Flex, Digital
EX2900000141	1 to 3	A	16 Feb 21	PCB Heat Spreader TLM (2x5)
EX2900000161	1 to 4	A	16 Feb 21	PCB Flex, PDMx1
EX2900000162	1 to 3	A	16 Feb 21	Schematic, PCB FLEX, PDMx1
EX2900000171	1 to 4	A	16 Feb 21	PCB, ISEM Shield

CSA Group Netherlands B.V.  
 Utrechtseweg 310, Building B42,  
 6812AR, Netherlands

# Certificate Annexe



Certificate Number: CSANe 20ATEX1197X

Equipment: J22 TDLAS Gas Analyzer, J22 TDLAS Gas Analyzer System with Sample Conditioning System on Panel, J22 TDLAS Gas Analyzer System with Heated Enclosed Sample Conditioning System

Applicant: SpectraSensors Inc.

Drawing	Sheets	Rev.	Date (Stamp)	Title
EX5500000004	1 to 2	B	16 Feb 21	Pressure Sensor, Digital
EX6000000009	1 of 1	A	16 Feb 21	Cable, P3 to ISE MCU Digital Brd
EX6000000012	1 of 1	A	16 Feb 21	Cable Assy, Pressure, Digital
EX8000000062	1 to 3	C	16 Feb 21	Isem Interface Module Assy
EX8000000092	1 to 3	E	16 Feb 21	Analyzer Assy, 0.8m, Unit Sealing
EX8000000093	1 to 4	C	16 Feb 21	Analyzer System Assy, Unit Sealing
EX8000000104	1 of 1	A	16 Feb 21	Cover Assy, Opt Block
EX8000000107	1 to 2	B	16 Feb 21	Det Module, Parabolic, TO18, Sig
EX8000000111	1 to 2	A	16 Feb 21	Optical Head Assy, Cert
EX8000000112	1 to 3	D	16 Feb 21	Analyzer Assy, Cert
EX8000000113	1 to 5	A	16 Feb 21	TLM ASSY, Cert
EX8000000114	1 to 11	A	16 Feb 21	SCS Assy, Enclosure, J22, Certification
EX8700000015	1 of 1	A	16 Feb 21	T-Plate Assy, 0.8m

CSA Group Netherlands B.V.  
Utrechtseweg 310, Building B42,  
6812AR, Netherlands